

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1, Dordrecht
Person responsible: Ing. C. Oosterman

Applicant

Name: Ecometros S.L.
Address: C/Urgel, 240 2°C
08036 Barcelona
Spain

Manufacturer of the certified type

Name: Ecometros S.L.
Address: C/Urgel, 240 2°C
08036 Barcelona
Spain

Identification of the certified type

Diaphragm gas meter, with the following characteristics:

The meter, without physical changes, can be classified into one of the following G-values:

model	Q_{\max} (m ³ /h)	Q_{\min} (dm ³ /h)	p_{\max} (Bar)	V (dm ³)
G4(S)/G4(A)	6	40	0,5 or 1	1,2
G2,5(S)/G2,5(A)	4	25	0,5 or 1	1,2
G1,6(S)/G1,6(A)	2,5	16	0,5 or 1	1,2



Nederlands Meetinstituut

OIML Member State
The Netherlands

OIML Certificate N° R31/1995-NL1-04.01 rev. 1

Project number 600113

Page 2 of 2

This Certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R31


Edition 1995 (E)

This Certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML Internal Recommendation(s).

This Certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated Test Reports:
N° R31/1995-NL1-04.01 that includes 16 pages;
N° R31/1995-NL1-04.01b that includes 8 pages.

The issuing authority
Ing. C. Oosterman
Manager Product Certification



24 January 2006

*
* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.