

OIML Certificate N° R31/1995-NL1-04.02 rev.01

Project number 503666 Page 1 of 2

OIML Member State The Netherlands

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name:

NMi Certin B.V.

Address:

Hugo de Grootplein 1, Dordrecht

Person responsible:

Ing. C. Oosterman

Applicant

Name:

DESTAS DIJITAL ELEKTRONIK SAN. TIC. A.S.

Address:

Seyhli Mah Mimarsinan Cd. Hilak Sk. No: 33

34906, Kurtkoy - Pendik - Istanbul

Turkey

Manufacturer of the certified type

Name:

DESTAS DIJITAL ELEKTRONIK SAN. TIC. A.S.

Address:

Sevhli Mah Mimarsinan Cd. Hilak Sk. No: 33

34906, Kurtkoy - Pendik - Istanbul

Turkey

Identification of the certified type

Diaphragm gas meter, with the following characteristics:

model	Q _{max}	Q _{min}	p _{max}	V
	(m³/h)	(dm³/h)	(bar)	(dm³)
DESGAM 4	6	40	0,5	1,2



Nederlands Meetinstituut

OIML Certificate N° R31/1995-NL1-04.02 rev.01

OIML Member State The Netherlands

Project number 503666 Page 2 of 2

This Certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R31

Edition 1995 (E)

This Certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML Internal Recommendation(s).

This Certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated Test Report: R31/1995-NL1-04.02 that includes 17 pages.

The issuing authority Ing. C. Oosterman

Manager Product Certification

12/January 2006

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.