## Physikalisch-Technische Bundesanstalt

### Braunschweig und Berlin

Member State of OIML Germany



OIML Certificate N° R49-1/2006-DE1-07.03 Revision 2

### OIML CERTIFICATE OF CONFORMITY

#### **Issuing Authority**

Name:	Physikalisch-Technische Bundesanstalt
Address:	Bundesallee 100, 38116 Braunschweig
Person responsible:	Dr. Gudrun Wendt

#### Applicant

Name:	Elster Metering Limited
	130 Camford Way

Address: Sundon Park Luton, Bedfordshire LU3 3AN LU3 3AN United Kingdom

Manufacturer of the certified type is the applicant.

Identification of the certified typeWater Meter intended for the metering of cold potable water Type: SM100VR, SM150VR
--

Further characteristics see page 3

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R49-1 (2006): Metrological and technical requirements R49-2 (2006): Test methods R49-3 (2006): Test report format

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

## **Physikalisch-Technische Bundesanstalt**

OIML Certificate N° R49-1/2006-DE1-07.03 Revision 2

The conformity was established by the results of tests and examinations provided in the associated Report No. PTB-1.5-4030627, Revision 1 (96 pages) and Test Report No. PTB-1.5-4036395, Revision 1 (93 pages).

The Issuing Authority

The CIML Member

Dr. Gudrun Wendt

Head of Department Liquid Flow

05.08.2009

Dr. Roman Schwartz

Head of Division Mechanics and Acoustics

05.08.2009

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.

# Physikalisch-Technische Bundesanstalt

OIML Certificate N° R49-1/2006-DE1-07.03 Revision 2

Identification of the certified pattern - page 1 continued

Metrology characteristics SM150VR:

Q <sub>3</sub> :	2.5 $m^{3}/h$
Q <sub>4</sub> :	3.125 $m^{3}/h$
Q <sub>2</sub> /Q <sub>1</sub> :	1.6
Q <sub>1</sub> :	0.0156 $m^{3}/h$ 0.0125 $m^{3}/h$
Q <sub>2</sub> :	0.025 $m^{3}/h$ 0.020 $m^{3}/h$
Q <sub>3</sub> /Q <sub>1</sub> :	160 200
Lenght:	110 mm
Thread:	G ¾" B
Measuring principle:	Fluidic oscillation
Accuracy Class:	2
Temperature Class:	T30
Maximum admissible pressure:	1,6 MPa (16 bar)
Environmental Class:	B and C
Maximum admissible temperature:	30 (°C)
Metrology characteristics SM100VR:	
$Q_3:$	1.6 $m^{3}/h$
$Q_4:$	2.000 $m^{3}/h$
$Q_2/Q_1:$	1.6
$Q_1:$	0.01 $m^{3}/h$ 0.008 $m^{3}/h$ 0.0064 $m^{3}/h$
$Q_2:$	0.016 $m^{3}/h$ 0.0128 $m^{3}/h$ 0.01024 $m^{3}/h$
$Q_3/Q_1:$	160 200 250
Lenght:	110 mm
Thread:	G ¾" B
Measuring principle:	Fluidic oscillation
Accuracy Class:	2
Temperature Class:	T30
Maximum admissible pressure:	1,6 MPa (16 bar)
Environmental Class:	B and C
Maximum admissible temperature:	30 (°C)
Installation details SM100VR and SM150VR:	
Connection type:	Screw thread
Minimum straight length of inlet pipe:	0 mm
Minimum straight length of outlet pipe:	0 mm
Flow conditioner:	none
Orientation limitations:	none