

Member State
Denmark

OIML Certificate N°
R51/1996-DK1-05.01

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name:

The Danish Accreditation and Metrology Fund

Address:

Dyregårdsvej 5B
DK-2740 Skovlunde
Denmark

Person responsible:

P. Claudi Johansen

Applicant

Name:

Kosan Crisplant A/S

Address:

P.O. Pedersens Vej 22
DK-8200 Aarhus N
Denmark

Manufacturer

of the certified pattern:

Kosan Crisplant A/S

Identification

of the certified pattern:

Automatic Checkweighing Instrument
Type: ECS Check Weigher,
ECS Filling Machine &
ECS Weight Correction Machine

Further characteristics are set out on page 2.

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R51
edition 1996
for accuracy class X(1) and Y(a)

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

**OIML Certificate N°
R51/1996-DK1-05.01**

The conformity was established by tests described in the associated test reports from DELTA, DK N° DANAK-199756, that includes 10 pages and 6 annexes (113 pages), and N° DANAK-199757, that includes 10 pages and 6 annexes (109 pages).

The issuing authority


The Danish Accreditation and Metrology Fund

Date: 07. MAR 2005

The CIML member

P. Claudi Johansen

Date: 07. MAR 2005



Characteristics:

Accuracy class		X(1) and Y(a)
Weighing range		Single-interval
Number of intervals	n	≤ 3000
Maximum capacity	Max	30, 60, 120 or 600 kg
Minimum capacity	Min	20e
Maximum conveyer speed		20 m/min
Connected analogue load cell		One or two, Flintec type PC 6 or PCB, 50, 100, 200, 400 or 1000 kg
Temperature range		-10 °C / +40 °C
Weighing principle		Semi-automatic (static) or non-automatic (static)

Important note:

Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.