

Nederlands Meetinstituut

**OIML Member state**  
The Netherlands

**OIML Certificate N° R51/1996-NL1-05.01 revision 1**  
Project number 505317  
Page 1 of 2

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: NMI Certin B.V.  
Address: Hugo de Grootplein 1,  
3314 EG Dordrecht  
The Netherlands  
Person responsible: Ing. C. Oosterman

### Applicant

Name: Anritsu Industrial Solutions Co. Ltd.  
Address: 1800 Onna Atsugi-shi  
Kanagawa-Prefecture, 243  
Japan

### Manufacturer of the certified type

Name: Anritsu Industrial Solutions Co. Ltd.  
Address: 1800 Onna Atsugi-shi  
Kanagawa-Prefecture, 243  
Japan

### Identification of certified type

Type : SV series

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T6483 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R51**  
Edition 1996 (E)  
for accuracy class X(1).

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.


**NMI Certin B.V.**  
Hugo de Grootplein 1, 3314 EG Dordrecht  
P.O. Box 394, 3300 AJ Dordrecht, NL  
phone +31 78 6332332  
fax +31 78 6332309  
certin@nmi.nl  
www.nmi.nl

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI B.V. (see "Regulation objection and appeal against decisions of NMI B.V.")

NMI Certin B.V., chamber o.c. nr. 27.233.418

This document is issued under the provision that no responsibility is accepted and that the applicant gives warranty for each responsibility against third parties.

The notification of NMI Certin as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org).



This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

- N° R51/1996-NL1-04.02A that includes 51 pages;
- N° R51/1996-NL1-04.02B that includes 41 pages;
- N° R51/1996-NL1-04.02C that includes 31 pages;
- N° R51/1996-NL1-05.01A that includes 21 pages;
- N° R51/1996-NL1-05.01B revision 1 that includes 21 pages.

This revision certificate replaces the earlier version.

The Issuing Authority  
Ing. C. Oosterman  
Manager Product Certification

8 September 2005

\*  
\* \*

Identification of the certified type (continued)

Weighcell	Electromagnetic balance				Load cell
	KW1062A		KW1064A		KW105..A..
Temperature range	0 °C / 40 °C	5 °C / 40 °C	0 °C / 40 °C	5 °C / 40 °C	0 °C / +40 °C
Max ≤	600 g		3000 g		25000 g
Min ≥	10 g	15 g	20 g	20 g	20 g
e ≥	0.5 g	0.2 g	1 g	0.5 g	0.5 g
n ≤	6000 for single interval instruments				
n ≤	3000 per partial weighing range for instruments with multi-interval. Maximum of 3 partial weighing ranges				
T ≤	-100% of Max				
Maximum belt speed	120 m/min	100 m/min	120 m/min	100 m/min	100 m/min
Maximum weighing speed packages per minute.	600	600	400	342	400

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.