



Member State of OIML  
Germany



OIML Certificate No.  
**R51/2006-DE1-15.02**

## OIML CERTIFICATE OF CONFORMITY

### Issuing Authority

Name: Physikalisch-Technische Bundesanstalt  
Address: Bundesallee 100, 38116 Braunschweig  
Person responsible: Dr. O. Mack

### Applicant

Name: Pfreundt GmbH  
Address: Robert-Bosch-Str. 5, 46354 Südlohn  
GERMANY

Manufacturer of the certified type is the applicant.

### Identification of the certified type

Automatic catchweigher, integrated into a hydraulic system of front-end loaders and other vehicles  
Type: WK60

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R51-1**, edition 2006  
for accuracy class Y(b),  $n \leq 300$

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.



**OIML Certificate No.**  
**R51/2006-DE1-15.02**

The conformity was established by the results of tests and examinations provided in the associated Report 1.12-4064672 and Test Reports

No. 1.12-4064672/1	that includes 16 pages
No. 1.12-4064672/2	that includes 16 pages
No. 1.12-4064672/3	that includes 8 pages
No. 1.12-4057260/1	that includes 34 pages
No. 1.12-4034525/1	that includes 21 pages
No. 1.12-4034527/2	that includes 17 pages
No. 1.12-00084041	that includes 19 pages

Further technical data: see Report 1.12-4064672.

**The Issuing Authority**

Dr. O. Mack  
Member of Certification Body

27.10.2015

**The OIML Member**

Dr. R. Schwartz  
Vice President

27.10.2015

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.