



# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R51/2006-NL1-12.01  
Project number 11200700  
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant	Ishida Europe Ltd. 11 Kettles Wood Drive Woodgate business Park Birmingham B32 3DB United Kingdom
Manufacturer	Ishida Europe Ltd. 11 Kettles Wood Drive Woodgate business Park Birmingham B32 3DB United Kingdom
Identification of the certified type	An <b>Automatic catchweighing instrument</b> Type : DACS-G-S015... and DACS-G-S060...
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R51** - Edition 2006 (E) for accuracy class XIII(1)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**  
30 March 2012

C. Oosterman  
Head Certification Board

NMi Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see [www.nmi.nl](http://www.nmi.nl)).





# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R51/2006-NL1-12.01  
Project number 11200700  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- Number R51/2006-NL1-10.03A dated 6 July 2010 that includes 40 pages, plus Annex that includes 30 pages;
- Number R51/2006-NL1-10.03B dated 6 July 2010 that includes 31 pages, plus Annex that includes 30 pages;
- Number R51/2006-NL1-10.03C dated 6 July 2010 that includes 20 pages.

## Characteristics of the automatic catchweighing instrument

- Max :  $600 \text{ g} \leq \text{Max} \leq 6000 \text{ g}$
- Min :  $\geq 15 \text{ g}$  maximum belt speed 60 m/min  
 $\geq 30 \text{ g}$  maximum belt speed 100 m/min
- Verification scale interval :  $e \geq 0,2 \text{ g}$
- Maximum number of scale intervals :  $n \leq 3000$  divisions per partial weighing range for instruments with multi-interval
- Maximum partial weighing ranges : 2
- Preset tare :  $\text{PT} \leq - \text{Max}$
- Maximum belt speed : 100 m/min.
- Temperature range :  $- 5 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$