



Member State of OIML
United Kingdom of Great Britain
and Northern Ireland

OIML Certificate No
R60/1991-GB-00.05
Revision 1

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: **National Weights and Measures Laboratory**
Address: **Stanton Avenue
Teddington
Middlesex
TW11 0JZ
United Kingdom**

Person responsible: **P R Dixon
Product Certification Manager**

Applicant

Name: **Lahti Precision Oy (formerly Raute Precision Oy)**
Address: **Mestarinkatu 10
15800 LAHTI
Finland**

Manufacturer of the certified pattern is the Applicant.

Identification of the certified pattern:

Model TB5
Further characteristics see page 2

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organisation of Legal Metrology -OIML):

OIML:	R60
Edition:	1991 (E)
Accuracy class:	C3
$p_{LC} = 0.7$	

OIML Certificate No
R60/1991-GB1-00.05
Revision 1

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.


The conformity was established by tests described in the associated test report: N° TR00338 that includes 24 pages and test report: N° TR00339 which includes 26 pages.

Issuing authority

CIML member



Mr P R Dixon



Dr J W Llewellyn

for NWML

Date 26th February 2007

Ref: T1136/0018

Maximum capacity E_{\max}	500, 1000, 2000, 5000 kg
Minimum dead load	0 kg
Accuracy Class	C3
Maximum number of load cell intervals (n)	3000
Minimum LC verification interval (V_{\min})	$E_{\max}/6250$

Maximum output	Input impedance	Output impedance	Cable type	Standard cable length
2 mV/V	400 \pm 20 Ω	350 \pm 3 Ω	6-wire dual floating screen	5 m

This certificate replaces the previous version of the certificate.

Important note: Apart from the mention of the certificates reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.