

Member State
P.R.China



OIML Certificate No.
R60/2000-CN1-08.14

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine
Address: NO.9 Madiandonglu, Haidian District, Beijing, 100088 China

Applicant

Name: ZHEJIANG SOUTH-OCEAN SENSOR MANUFACTURING CO., LTD.
Address: NO. 888, XINGYUAN STREET, QIANLONG DEVELOPMENT ZONE,
QIANYUAN TOWN, DEQING COUNTY, ZHEJIANG PROVINCE, (313216) CHINA

Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:

Address:

Identification of the certified type

Load Cell

Model designation	GX NH
Accuracy class	C
Maximum capacity (E_{max})	1000kg, 1500kg, 2000kg, 2500kg, 3000kg, 5000kg ;
Minimum load cell verification interval (V_{min})	150g, 225g, 300g, 375g, 450g, 750g ;
Minimum dead load (E_{min})	10kg, 20kg, 30kg, 30kg, 40kg, 60kg
Maximum number of load cell interval (n_{max})	3000
Apportionment factor (P_{Lc})	0.7
Safe overload	150% E_{max}
Operating Temperature range	-10 ~ +40

The numbers of the associated test reports and the included pages:
LSfc2008-6005 20 pages

This certificate attests the conformity of the above-mentioned type(represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: 60

Edition: 2000

For accuracy class (if applicable): C3

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned ,as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval. The conformity was established by tests described in the associated test reports.

The issuing authority

Office of OIML Affairs

中国秘书处

December 24,, 2008

The CIML member

蒲长城

December 24,,2008

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.