

Member State  
P.R.China

OIML Certificate No.  
R60/2000-CN1-16.01



# OIML CERTIFICATE OF CONFORMITY

## Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine  
Address: N0.9 Madiandonglu, Haidian District, Beijing, 100088 China

## Applicant

Name: Mettler Toledo (ChangZhou) Precision Instrument Ltd.  
Address: 5# middle huashan road, Changzhou, Jiangsu, (213125) P.R.China

## Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:  
Address:

## Identification of the certified type

Load Cell

Model designation	SLB615D
Maximum capacity ( $E_{max}$ )	(220、550、1100、2200、4400) kg
Minimum dead load ( $E_{min}$ )	0 kg
Accuracy class	C
Maximum number of load cell interval ( $n_{max}$ )	$n_{max} \leq 10000$
Minimum load cell verification interval ( $V_{min}$ ) or Y	$v_{min} = 0.01$ kg
Apportionment factor ( $P_{LC}$ )	0.8
Safe overload	$150\% E_{max}$
Operating temperature range	$-10^{\circ}\text{C} \sim +40^{\circ}\text{C}$
Humidity classification	CH

The numbers of the associated test reports and the included pages:  
LSfc2016-6001 39 pages

This certificate attests the conformity of the above-mentioned type(represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: 60

Edition: 2000

For accuracy class (if applicable): C

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned ,as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval. The conformity was established by tests described in the associated test reports.

The issuing authority

  
*Office of OIML Affairs*

中国秘书处

**October 30, 2016**

The CIML member



**October 30, 2016**

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.