

Physikalisch-Technische Bundesanstalt

Braunschweig und Berlin

Member State of OIML
Germany



OIML Certificate N°
R60/2000-DE1-08.09

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: Physikalisch-Technische Bundesanstalt
Address: Bundesallee 100, 38116 Braunschweig
Person responsible: Dr. Panagiotis Zervos

Applicant

Name: Sartorius Mechatronics T&H GmbH
Address: Meiendorfer Straße 205, 22145 Hamburg
Germany

Manufacturer of the certified type is the applicant.

Identification of the certified type

Load Cell
Strain gauge double bending beam load cell

Type: 011xxK

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R60, edition 2000
for accuracy classes C3 ; C4 ; C6

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

Physikalisch-Technische Bundesanstalt

OIML Certificate N°
R60/2000-DE1-08.09

The conformity was established by the results of tests and examinations provided in the associated Test Reports

No. PTB 1.12-4037198-1 that includes 22 pages
 No. PTB 1.12-4037198-2 that includes 18 pages

The Issuing Authority

The OIML Member

Dr. P. Zervos
 Direktor und Professor

Dr. R. Schwartz
 Direktor und Professor

19.09.2008

19.09.2008

The load cells (LC) of the series 011xxK are double bending beam load cells made of stainless steel. The strain gauge application is hermetically sealed.

The metrological characteristics for application in approved weighing instruments are listed in table 1

Table 1: Essential data

Accuracy class			C3	C4	C6
Maximum number of load cell intervals		n_{LC}	3000	4000	6000
Rated output		mV/V	1,94		
range 1	Maximum capacity	E_{max}	kg	220 / 1760 / 2200 / 4400	
	Minimum load cell verification interval	v_{min}	$\% \cdot E_{max}$	0,0100	
range 2	Maximum capacity	E_{max}	kg	550 / 1100	
	Minimum load cell verification interval	v_{min}	$\% \cdot E_{max}$	0,0090	

Dead load: $0\% \cdot E_{max}$; Safe overload: $150\% \cdot E_{max}$; Input impedance: $> 350 \Omega$; Fraction: $p_{LC} = 0.7$

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Reports is not permitted, although either may be reproduced in full.