

# Physikalisch-Technische Bundesanstalt

Braunschweig und Berlin

Member State of OIML  
Germany



OIML Certificate No.  
**R60/2000-DE1-10.11**

## OIML CERTIFICATE OF CONFORMITY

### Issuing Authority

Name: Physikalisch-Technische Bundesanstalt  
Address: Bundesallee 100, 38116 Braunschweig  
Person responsible: Dr. Dirk Ratschko

### Applicant

Name: Zemic Europe B.V.  
Address: Leerlooierstraat 8  
4871 EN Etten-Leur  
Niederlande

Manufacturer of the certified type is the applicant.

### Identification of the certified type

Strain gauge single point load cell  
Type: L6T  
Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R60**, edition 2000  
for accuracy classes C3, C4

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

# Physikalisch-Technische Bundesanstalt

OIML Certificate No.  
**R60/2000-DE1-10.11**

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Report and Test Reports

Test-Report No. 1.12-4044287	that includes 22 pages
Test-Report No. 1.12-4044287-4	that includes 22 pages
Report No. 1.12-4044287	that includes 6 pages

## The Issuing Authority

## The CIML Member

Dr. D. Ratschko  
 Head of Department

Dr. R. Schwartz  
 Head of Division

18.10.2010

18.10.2010

The load cells of the series L6T are single point load cells. They are made of aluminium and the strain gauge application is hermetically sealed.

The metrological characteristics for application in approved weighing instruments are listed in table 1.

Table 1: Essential data

Accuracy class			C3	C4
Maximum number of load cell intervals	$n_{LC}$		3000	4000
Rated output		mV/V	2	
Maximum capacity	$E_{max}$	kg	50 / 75 / 100 / 150 / 200 / 250	300 / 500 / 635 / 1000
Minimum load cell verification interval	$\frac{V_{min}}{(E_{max} / Y)}$		$E_{max} / 6500$	$E_{max} / 15000$
				$E_{max} / 6500$

Dead load:  $0\% \cdot E_{max}$ ; Safe overload:  $150\% \cdot E_{max}$ ; Input impedance: 409  $\Omega$ ; Fraction:  $p_{LC} = 0.7$

**Important note:** Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Reports is not permitted, although either may be reproduced in full.