



Member State of OIML
United Kingdom of Great Britain
and Northern Ireland

OIML Certificate No
R60/2000-GB1-07.02

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: **National Weights and Measures Laboratory**
Address: **Stanton Avenue
Teddington
Middlesex
TW11 0JZ
United Kingdom**

Person responsible:

Paul Dixon – Product Certification Manager

Applicant

Name: **Vehicle Weighing Solutions Ltd**
Address: **Hyde Road
Off Foxdenton Lane
Chadderton
Oldham
M24 1QG
United Kingdom**

Manufacturer of the certified pattern is:

The applicant

Identification of the certified pattern:

**Stainless steel single ended shear beam (bending) strain
gauge load cell**

Model Designation	VWS850
Maximum capacity, E_{\max}	500, 1000, 1500, 2000, 2500 kg
Accuracy class	C1.5
Maximum number of load cell intervals, n_{\max}	1500
Minimum verification interval, V_{\min}	$E_{\max}/5000$
Apportionment factor; p_{LC}	0.7

OIML Certificate No
R60/2000-GB1-07.02

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organisation of Legal Metrology -OIML):

R 60 *Metrological regulation for load cells* Edition: 2000 (E) for accuracy class: C1.5

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated NWML test report, N° TR: 515 which includes 23 pages.

Issuing authority

CIML member




Mr P R Dixon
for NWML

Dr J W Llewellyn

Date 30 May 2007
Ref: T1136/0011

Table 1: Essential technical data

<i>Model designation</i>	<i>Designation</i>	<i>Value</i>	<i>Units</i>
Classification		C1.5	
Additional marking		CH	
Maximum number of load cell verification intervals	n_{LC}	1500	
Maximum capacity	E_{max}	500, 1000, 1500, 2000, 2500	kg
Minimum dead load, relative	E_{min}/E_{max}	0	%
Relative V_{min} (ratio to minimum LC verification interval)	$Y = E_{max}/V_{min}$	5000	
Relative DR (ratio to minimum dead load output return)	$Z = E_{max}/(2*DR)$	12500	
Rated output		2.2	
Maximum excitation voltage		15	V dc
Input impedance (for strain gauge LCs)	R_{LC}	400	Ω
Temperature rating		-10/+40	$^{\circ}C$
Safe overload, relative	E_{lim}/E_{max}	150	%
Cable length		6	m
Additional characteristics		4 wire + screen	

Important note: Apart from the mention of the certificates reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.