



Member State of OIML
Japan



OIML Certificate No.
R60/2000-JP1-12.02

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: National Metrology Institute of Japan / National Institute of
Advanced Industrial Science and Technology (NMIJ / AIST)
Address: AIST Tsukuba Central 3-9, Tsukuba Ibaraki 305-8563, Japan
Person responsible: Dr. Tamotsu Nomakuchi, President of AIST

Applicant

Name: A&D Company, Limited
Address: 3-23-14, Higashi-Ikebukuro, Toshima-ku, Tokyo 170-0013, Japan

Manufacturer of the certified pattern

Name: A&D Company, Limited
Address: 3-23-14, Higashi-Ikebukuro, Toshima-ku, Tokyo 170-0013, Japan

Identification of the certified pattern:

Tension load cell
Type: LC1205- T001A , LC1205- T002, LC1205- T005
Fraction: $\pi=0.7$
Temperature range: $-10\text{ }^{\circ}\text{C} / 40\text{ }^{\circ}\text{C}$



Member State of OIML
Japan



OIML Certificate No.
R60/2000-JP1-12.02

Characteristics:

Model designation			LC1205-T001A	LC1205-T002	LC1205-T005
Accuracy class	Class	-	C		
Maximum number of load cell verification intervals	n_{max}	-	4000 3000		
Humidity symbol			CH		
Minimum dead load	E_{min}	kg	0		
Maximum capacity	E_{max}	kg	1000	2000	5000
Safe load limit	E_{lim}	kg	$1.5 * E_{max}$		
Minimum verification interval	v_{min}	kg	$E_{max} / 10000$		
Apportionment factor	p_{LC}		0.7		
Ratio of minimum LC Verification interval $Y = E_{max} / v_{min}$	Y	-	10000		
Ratio of minimum dead load output return $Z = E_{max} / (2 * DR)$	Z	-	4000 in the case of $n_{max} = 4000$		
Rated output		mV/V	2		
Maximum excitation voltage		V AC/DC	15		
Input impedance	R_{LC}	Ω	440		
Cable detail		-	5m 4wire		

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report(s) with the requirements of the following Recommendation of the International Organization of Legal Metrology - OIML):

R60, edition 2000 (E)
For accuracy class C

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test report no. 12-05/R60:2000, that includes 19 pages.



Member State of OIML
Japan



OIML Certificate No.
R60/2000-JP1-12.02

The Issuing Authority
NMIJ/AIST



Dr. T. Nomakuchi
President of AIST
2012-03-09

The OIML member

Dr. Y. Miki
2012-03-09

Important note: Apart from the mention of certificate's reference number and the name of the OIML Member State in which the certificate is issued, partial quotation of the certificate and of the associated Test Report is not permitted, although either may be reproduced in full.