



**OIML Member State**  
The Netherlands

**OIML Certificate N° R60/2000-NL1-05.01**  
Project number 501482  
Page 1 of 2

## OIML CERTIFICATE OF CONFORMITY

### Issuing Authority

Name: NMi Certin B.V.  
Address: Hugo de Grootplein 1, Dordrecht  
Person responsible: Ing. C. Oosterman

### Applicant

Name: Bongshin Loadcell Co., Ltd.  
Address: 148, Sangdaewon-dong, Jungwon-ku,  
Seongnam-city, Geonggi-do,  
Korea


### Manufacturer of the certified type

Name: Bongshin Loadcell Co., Ltd.  
Address: 148, Sangdaewon-dong, Jungwon-ku,  
Seongnam-city, Geonggi-do,  
Korea

### Identification of the certified type

Type : DBBP Series  
Fraction :  $P_i = 0.7$   
Temperature range  $-10\text{ }^{\circ}\text{C} / 40\text{ }^{\circ}\text{C}$

Maximum capacity ( $E_{\max}$ )	200 kg up to and including 1000 kg
Accuracy class	C
Maximum number of load cell verification intervals ( $n_{\max}$ )	3000
Ratio of minimum LC verification interval $Y = E_{\max} / V_{\min}$	3333
Ratio of minimum dead load output return $Z = E_{\max} / (2 * DR)$	4600



Nederlands Meetinstituut

**OIML Certificate N° R60/2000-NL1-05.01**

Project number 501482

Page 2 of 2

**OIML Member State**

The Netherlands

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the test certificate and the description with number TC6726 and the appertaining documentation folder), with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R60

Edition 2000 (E)

for accuracy class C

This Certificate relates only to the metrological and technical characteristics of the type of the instrument covered by the relevant OIML International Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° R60/2000-NL1-05.01 that includes 40 pages.

The Issuing Authority  
Ing. C. Oosterman  
Manager Product Certification

*D/a*

16 March 2005

The OIML Member  
Dr. ir. C. van Mullem

16 March 2005

\*  
\* \*

---

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.