



OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Mettler-Toledo (Changzhou) Precision Instrument Ltd.
Address: No. 5 Middel HuaShan Road
Changzhou, Jiang Su
P.R. of China


Manufacturer of the certified type

Name: Mettler-Toledo (Changzhou) Precision Instrument Ltd.
Address: No. 5 Middel HuaShan Road
Changzhou, Jiang Su
P.R. of China

Identification of certified type

Single point, bending beam load cell
Type : MT1241-...
Fraction : $P_i = 0.7$
Temperature range: -10 °C / +40 °C

Maximum capacity (E_{max})	30 kg up to and including 250 kg
Accuracy Class	C
Maximum number of load cell intervals (n)	3500
Ratio of minimum LC Verification interval $Y = E_{max} / V_{min}$	6000



Nederlands Meetinstituut

OIML Member state
The Netherlands

OIML Certificate N° R60/2000-NL1-06.04

Project number 603081

Page 2 of 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the test certificate and the description with number TC5382 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R60
Edition 2000 (E)
for accuracy class C

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

N° R60/2000-NL1-06.04A that includes 37 pages;

N° R60/2000-NL1-06.04B that includes 38 pages.

The Issuing Authority
Ing. C. Oosterman
Manager Product Certification


29 March 2006

*
* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.