

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.

Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

Person responsible: Ing. C. Oosterman

Applicant

Name: Dini Argeo S.r.l.

Address: Via della Fisica 20
41040 - Spezzano di Fiorano (MO)
Italy

Manufacturer of the certified type

Name: Dini Argeo S.r.l.

Address: Via della Fisica 20
41040 - Spezzano di Fiorano (MO)
Italy

Identification of certified type

A digital compression load cell

Type : RCD ...

Fraction : $P_i = 0.8$

Temperature range $-10\text{ °C} / +40\text{ °C}$

Maximum capacity (E_{max})	15 t up to and including 75 t			
Accuracy Class	C			
Maximum number of load cell intervals (n)	1000	3000	4000	5000
Ratio of minimum LC Verification interval $Y = E_{max} / V_{min}$	15 000			
Ratio of minimum dead load output return $Z = E_{max} / (2 * DR)$	8000			

OIML Member state

The Netherlands

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the test certificate and the description with number TC7078 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R60

Edition 2000 (E)
for accuracy class C

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

N° R60/2000-NL1-06.09A that includes 55 pages;

N° R60/2000-NL1-06.09B that includes 13 pages.

The Issuing Authority NL1
NMI Certin, 26 March 2009



C. Oosterman
Head Certification Board

*

* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.