

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Dini Argeo S.r.l.
Address: Via della Fisica 20
41040 - Spezzano di Fiorano (MO)
Italy

Manufacturer of the certified type

Name: Dini Argeo S.r.l.
Address: Via della Fisica 20
41040 - Spezzano di Fiorano (MO)
Italy

Identification of certified type

A compression load cell
Family of type : RCA ...
Fraction : $P_i = 0.7$
Temperature range : $-10\text{ °C} / +40\text{ °C}$

Maximum capacity (E_{max})	5 t up to and including 100 t
Accuracy Class	C
Maximum number of load cell intervals (n)	4000
Ratio of minimum LC Verification interval $Y = E_{max} / V_{min}$	15000
Ratio of minimum dead load output return $Z = E_{max} / (2 * DR)$	4100



OIML Member state

The Netherlands

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the test certificate and the description with number TC5863 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R60

Edition 2000 (E)
for accuracy class C

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

- N° R60/2000-NL-01.08A, that includes 37 pages;
- N° R60/2000-NL-01.08B, that includes 40 pages;
- N° R60/2000-NL1-05.22, that includes 13 pages;
- N° R60/2000-NL1-07.10A, that includes 37 pages;
- N° R60/2000-NL1-07.10B, that includes 40 pages;
- N° R60/2000-NL1-07.10C, that includes 13 pages.

The Issuing Authority NL1
NMI Certin, 26 March 2009



C. Oosterman
Head Certification Board

*
* *