OIML Certificate No. R76/1992 - CN1-05.03

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: Office of OlML Affairs, General Administration of Quality Supervision, Inspection and Quarantine Address: N0.9 Madiandonglu, Haidian District, Beijing, 100088 China

Applicant

Name: , Changzhou Honsta electric manufacturing Co.,Ltd

Address: West end of zhongliangting flyover changxi Road, Changzhou, Jiangsu, 213001 P.R.China

Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:

Address:

Identification of the certified type

Price computing instrument

Type	ACS-15-JJ (F902-15)
Accuracy class	
Max	15kg
Min	100g
e	5g
n	3000g
Temperature range	0-40
Humidity range	85%RH

The numbers of the associated test reports and the included pages: 2004-57(0) 52pages

This certificate attests the conformity of the above-mentioned type(represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: 76

Edition: 1992

For accuracy class (if applicable):

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned ,as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval. The conformity was established by tests described in the associated test reports.

The issuit that the interior of the issuit that the interior of the interior

The CIML member



October 25, 2005

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.