

Physikalisch-Technische Bundesanstalt

Braunschweig und Berlin

Member State of OIML
Germany



OIML Certificate N°
R76/1992-DE1-07.04

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: Physikalisch-Technische Bundesanstalt
Address: Bundesallee 100, 38116 Braunschweig
Person responsible: Dr. Panagiotis Zervos

Applicant

Name: Mettler-Toledo GmbH
Address: Im Langacher, 8606 Greifensee
Switzerland

Manufacturer of the certified type is the applicant.

Identification of the certified type

Non-automatic electromechanical weighing instrument with or without lever works

Type: Classic SG-A

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76-1, edition 1992, including Amendment 1 (1994),
for accuracy class **II**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

Physikalisch-Technische Bundesanstalt

OIML Certificate N°
R76/1992-DE1-07.04

The conformity was established by the results of tests and examinations provided in the associated Report No. 1.12-4029743 (7 pages) and the Test Reports
No. 1.12-4029743/1 that includes 47 pages
No. 1.12-4029743/2 that includes 50 pages

The Issuing Authority

Dr. P. Zervos
Regierungsdirektor

2007-05-22

The OIML Member

Dr. R. Schwartz
Direktor und Professor

2007-05-22

Identification of the type (continued):

Compact weighing instrument with built-in display, keyboard and data-interface.

The weighing ranges comprising Max, verification scale intervals, number of verification scale intervals and scale intervals may be selected considering the limiting values in table 1.

Table 1

accuracy class	II
Max	610 g ... 6100 g
verification scale interval e	0.1 g ... 1 g
scale interval d	0.01 g ... 0.1 g
number of verification scale intervals n	≤ 6100
tare-balancing range	≤ 100 % of Max
temperature range	+10 °C ... +30 °C

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.