Physikalisch-Technische Bundesanstalt

Braunschweig und Berlin

Member State of OIML Germany



OIML Certificate No. R76/1992-DE1-07.07 Revision 1

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name:	Physikalisch-Technische Bundesanstalt
Address:	Bundesallee 100, 38116 Braunschweig
Person responsible:	Dr. Dirk Ratschko

Applicant

Name:	Schenck Process GmbH	
Address:	Pallaswiesenstr. 100, 64293 Darmstadt GERMANY	

Manufacturer of the certified type is the applicant.

Identification of the cer-	Non automatic electromechanical weighing instrument with or without
tified type	lever system
••	Type: DISOMAT Tersus

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76-1, edition 1992, including Amendment 1 (1994), for accuracy classes (III) (III)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

Physikalisch-Technische Bundesanstalt

OIML Certificate No. R76/1992-DE1-07.07 Revision 1

The conformity was established by the results of tests and examinations provided in the Report No. 1.12-4051368 that includes 13 pages and in the additional Test Report No. 1.12-4051368/1 that includes 9 pages.

The Issuing Authority

The CIML Member

Dr. D. Ratschko Head of Department Dr. R. Schwartz Head of Division

02.05.2011

02.05.2011

Identification of the type (continued):

Designed as weighbridge, road vehicle scale, rail-weighbridge, hopper scale, crane scale, overhead scale track as well as mobile instrument in conveyor systems. The weighing instrument is equipped with a lever system or directly introduces the force into one or more load cells.

The weighing ranges comprising Max, verification scale intervals, number of verification scale intervals and scale intervals may be selected considering the limiting values in table 1.

Table 1:

accuracy class		
Max	2 kg 600 t	2 kg 600 t
n ≤ ¹⁾	8000	1000
partial range $n_i \leq 2^{(i)}$	4000	1000
²⁾ Max / $e_1 \le$ or ³⁾ Max _r / $e_1 \le$	15000	5000
tare-balancing range	100 % of Max	
preset tare range	100 % of Max	1)
	100 % of Max ₁	2)
temperature range	-30 °C / +40 °C	4)

 $\stackrel{(1)}{\sim}$ this applies to each range of single- and multiple range instruments

²⁾ this applies only to multi-interval instruments

³⁾ this applies only to multiple range instruments

⁴⁾ in case the load cells are suitable for this temperature range.

n_i = Number of scale intervals for each partial weighing range

e₁ = Scale interval of the lowest partial weighing range

 $Max_r = Max$ of the largest range for a multiple weighing range instrument

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.