

Member State  
Denmark

**OIML Certificate N°  
R76/1992-DK1-09.05**

## **OIML CERTIFICATE OF CONFORMITY**

### **Issuing authority**

Name: **The Danish Accreditation and Metrology Fund**  
Address: Dyregårdsvej 5B  
2740 Skovlunde  
Denmark

Person responsible: P. Claudi Johansen

### **Applicant**

Name: **Taiwan Scale Mfg. Co., Ltd.**  
Address: 282, Sec. 3, Hoping W. Road  
Taipei  
TAIWAN

### **Manufacturer**

of the certified pattern: **Taiwan Scale Mfg. Co., Ltd.**

### **Identification**

of the certified pattern: **Non-automatic weighing instrument**  
**Type: T28**  
Further characteristics are set out on page 2.

This certificate attests the conformity of the above mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76**  
**edition 1992, including amendment 1 (1994)**  
**for accuracy class III**

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

Page 1. This certificate includes 2 pages

# OIML Certificate N° R76/1992-DK1-09.05

The conformity was established by tests described in the associated test reports from DELTA, DK, No. DANAK-1910465, dated 1 Apr. 2009, including 45 pages, DELTA, DK, No. DANAK-1910594, dated 14 Aug. 2009, including 36 pages.

## The issuing authority

The Danish Accreditation and Metrology Fund

Date: 16. Dec. 2009

## The OIML member

P. Claus Johansen

Date: 16. Dec. 2009



## Characteristics

Accuracy class		III
Weighing range		Single-interval or multi-interval (2 intervals)
Maximum capacity	Max	3 kg $\leq$ Max $\leq$ 25 kg
Minimum capacity	Min	20 e
Verification scale interval	e	$\geq$ 1 kg
Number of intervals	Max / e	$\leq$ 3000 (per interval)
Tare	T	$\leq$ Max
Temperature range for the indicator		0 °C / +40 °C

## Important note:

Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.