



Member State
Denmark

OIML Certificate N°
R76/1992-DK3-12.01

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: **DELTA**
Address: Venlighedsvej 4
2970 Hørsholm
Denmark

Person responsible: J. Hovgård Jensen

Applicant

Name: **TScale Electronics Mfg. (Kunshan) Co., Ltd.**
Address: No. 268 Zhujiawan Road,
Zhoushi, Kunshan, Jiangsu
CHINA

Manufacturer

of the certified pattern: **TScale Electronics Mfg. (Kunshan) Co., Ltd.**

Identification

of the certified pattern: **Non-automatic weighing instrument**
Type: TPS / TPSS / TPT Truck Pallet Scale
Further characteristics are set out on page 2.

This certificate attests the conformity of the above mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76
edition 1992, including amendment 1 (1994)
for accuracy class III and IIII

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

Page 1. This certificate includes 2 pages



OIML Certificate N° R76/1992-DK3-12.01

The conformity was established by tests described in the associated test reports from NMI, NL, No. 709226, dated 2008-04-17, including 44 pages, DELTA, DK, No. DANAK-1910424, dated 25 Feb. 2009, including 19 pages.

The issuing authority: **DELTA, OIML Issuing Authority DK3**
31 May 2012

J. Hovgård Jensen
Certification Officer

Characteristics

Accuracy class		III or IIII
Weighing range		Single-interval
Maximum capacity	Max	≤ 2000 kg
Minimum capacity	Min	20 e
Verification scale interval	e	≥ 1 kg
Number of intervals	Max / e	≤ 2000 (class III)
Tare	T	≤ -Max
Load receptor		Movable U-shaped pallet lifter
Temperature range for the indicator		-10 °C / +40 °C

Important note:

Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.