



Member State
Denmark

OIML Certificate N°
R76/1992-DK3-17.02

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: **DELTA**
Address: Venlighedsvej 4
2970 Hørsholm
Denmark

Person responsible: J. Hovgaard Jensen

Applicant

Name: **Tüm Elektronik Mühendislik San. ve Tic. Ltd.Şti.**
Address: İstanbul Deri Organize Sanayi 1. Yol, H7 Parsel,
Orhanlı
34956 Tuzla, İstanbul
TURKEY

Manufacturer

of the certified pattern: **Tüm Elektronik Mühendislik San. ve Tic. Ltd.Şti.**

Identification

of the certified pattern: **Non-automatic weighing instrument**
Type: EGE-LED / EGE-M
Further characteristics are set out on page 2.

This certificate attests the conformity of the above mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76
edition 1992, including amendment 1 (1994)
for accuracy class III

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

Page 1. This certificate includes 2 pages

DELTA



OIML Certificate N° R76/1992-DK3-17.02

The conformity was established by tests described in the associated test reports from DELTA, DK, No. DANAK-199788, dated 2005-12-21, including 101 pages.

The issuing authority: **DELTA, OIML Issuing Authority DK3**
25 April 2017

J. Hovgaard Jensen
Certification Officer

Characteristics

Accuracy class		III
Weighing range		Single-interval or 2 multi-ranges
Maximum capacity	Max	3 kg to 150 kg
Minimum capacity	Min	20 e
Verification scale interval	e	≥ 1 g
Number of intervals	n	≤ 3000 (per weighing range)
Temperature range for the indicator		-10 °C / +40 °C
Software version		31xzzz where x = 1, 2 or 4 and z = 0 to 9

Important note:

Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.