



Member State of OIML
United Kingdom of Great Britain
and Northern Ireland

OIML Certificate No
R76/1992-GB1-07.01

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: **National Weights and Measures Laboratory**
Address: **Stanton Avenue**
Teddington
Middlesex
TW11 0JZ
United Kingdom

Person responsible: **Paul Dixon**
Product Certification Manager

Applicant

Name: **Fabricantes de Basculas Torrey, S.A. de C.V.**
Address: **Ave. Los Andes 605**
Col. Coyoacan C.P. 64510
Monterrey NL
Mexico

Manufacturer of the certified pattern is the Applicant.

Identification of the certified pattern:

Torrey LSQ-20 and LSQ-40
Further characteristics see page 2

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML:	R76
Edition:	1992 (E)
Accuracy class:	III

OIML Certificate No
R76/1992-GB1-07.01

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.


The conformity was established by tests described in the associated test reports: SN: 00992 which includes 45 pages and checklist G20084 which includes 12 pages.

The issuing authority



Paul Dixon

The OIML member



Jeff Llewellyn

Date: 26 February 2007

Characteristics: Class III non-automatic weighing instrument designated the Torrey LSQ-20 and LSQ-40.

Instrument	LSQ-20	LSQ-40
Maximum capacity	20 kg	40 kg
Minimum load	100 g	200 g
Verification scale interval	5 g	10 g
Operating temperature range	-10 °C to + 40 °C	
Power supply	230 V ac 50 Hz	
Load cell	Medidata model M022	

Important note: Apart from the mention of the certificates reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.