



**Member State of OIML  
United Kingdom of Great Britain  
and Northern Ireland**

**OIML Certificate No  
R76/1992-GB1-07.03**

## **OIML CERTIFICATE OF CONFORMITY**

Issuing authority

Name: **National Weights and Measures Laboratory**  
Address: **Stanton Avenue  
Teddington  
Middlesex  
TW11 0JZ  
United Kingdom**

Person responsible: **Paul Dixon  
Product Certification Manager**

Applicant

Name: **NCR Corporation**  
Address: **2651 Satellite Blvd  
Duluth  
Georgia 30096  
USA**

Manufacturer of the certified pattern is the Applicant.

Identification of the certified pattern:

**NCR 7878-2000 non-automatic weighing instrument  
Further characteristics see page 2**

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

|                        |                 |
|------------------------|-----------------|
| <b>OIML:</b>           | <b>R76</b>      |
| <b>Edition:</b>        | <b>1992 (E)</b> |
| <b>Accuracy class:</b> | <b>III</b>      |

**OIML Certificate No  
R76/1992-GB1-07.03**

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.


The conformity was established by tests described in the associated test report Number SN:00997 having 46 pages, and associated pattern evaluation checklist G20070 having 12 pages.

The issuing authority



Mr P R Dixon

The OIML member



Dr J W Llewellyn

Date: 19 March 2007

Ref: T1128/0064

Characteristics: Mains powered, Class III, scanner/scale, non-automatic weighing instrument designated NCR 7878-2000

Max 15.000 kg  
Min 0.100 kg  
e = 0.005 kg

- Designed for flush-mounting in checkout surface.
- Steel framework supporting a barcode scanner, electronics, and a Flintec PBNS-30 scale module containing four loadcells.
- NCR 7825 remote display unit.
- NCR 7878 barcode scanner.
- Interface for POS connection.
- 90-240 Vac, 50/60 Hz supply (remote power supply module).
- Temperature range 0°C to +40°C

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.