



**Member State of OIML
United Kingdom of Great Britain
and Northern Ireland**

**OIML Certificate No
R76/1992-GB1-07.11**

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: **National Weights and Measures Laboratory**
Address: **Stanton Avenue
Teddington
Middlesex
TW11 0JZ
United Kingdom**

Person responsible: **Paul Dixon
Product Certification Manager**

Applicant

Name: **Avery Berkel**
Address: **Foundry Lane
Smethwick
West Midlands
B67 9DF
United Kingdom**

Manufacturer of the certified pattern is the Applicant.

Identification of the certified pattern:

**FX50 non-automatic weighing instrument
Further characteristics see page 2**

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML:	R76
Edition:	1992 (E)
Accuracy class:	III

**OIML Certificate No
R76/1992-GB1-07.11**

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test reports:

00193 having 16 pages (Avery Berkel)

00195 having 48 pages (Avery Berkel)

00196 having 10 pages (Avery Berkel)

SN: 1032 having 9 pages (NWML)

and associated pattern evaluation checklist G20136 having 12 pages.

The issuing authority



Mr P R Dixon

The OIML member



Mr P Mason

Date: 19 October 2007

Ref: T1128/0121

Characteristics: The Avery Berkel FX 50 is a Class III, self indicating, price computing, mains or battery powered non-automatic weighing instrument.

15 kg ≤ Maximum capacity ≤ 30 kg

100 g ≤ Minimum capacity ≤ 200 g

e ≥ 5 g

n ≤ 3000 for Class III instruments with single interval

The instrument is provided with the following devices:

Initial zero setting device (< 20 % of Max.)

Semi-automatic zero setting device (< 4 % of Max.)

Zero-tracking device

Zero indicator

Net indicator

Semi-automatic subtractive tare balancing device

Price computing device

Stability indicator

Low battery indicator

Power on/charging indicator

Fix indicator

100 g / kg indicators

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.