

Member State of OIML **United Kingdom of Great Britain** and Northern Ireland

OIML Certificate No R76/1992-GB1-09.03

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: **National Weights and Measures Laboratory**

Stanton Avenue Address:

Teddington Middlesex TW11 0JZ

United Kingdom

Paul Dixon Person responsible:

Product Certification Manager

Applicant

Name: Adam Equipment Co. Ltd.

Bond Avenue Address:

Bletchley

Milton Keynes MK1 1SW

United Kingdom

Identification of the certified pattern:

GFK..M non-automatic weighing instrument formed by connecting the GK indicator to a weighing platform

Further characteristics see page 2

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

> OIML: **R76** 1992 (E) **Edition: Accuracy class:** Ш

OIML Certificate No R76/1992-GB1-09.03

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated:

Test report: TR: 543 having 37 pages Pattern evaluation checklist: F20351 having 12 pages

The issuing authority

The CIML member

Mr P R Dixon Mr P Mason

Date: 13 March 2009 Ref: T1127/0020

Characteristics: The GFK..M instrument utilises the digital indicating device designated the

GK indicator connected to a weighing platform to form a Class III, mains or

battery-powered self-indicating, non-automatic weighing instrument.

Main features:

- Main PCB with processor type ATMEGA64
- 6-digit (plus symbols and indicators) LCD screen, and 9 LEDs to its left
- 20-key keypad (numerical and functions)
- Plastic enclosure

Devices:

- Initial zero setting
- Semi-automatic zero setting
- Zero tracking
- Zero indicator
- Semi-automatic subtractive tare weighing
- Net indicator
- Determination of stability of equilibrium
- Indication of stability of equilibrium
- Parts counting
- Percent weighing
- Checkweighing
- Live animal weighing
- Accumulation
- Real-time clock

Load cell:

The load cell may be:

| MODEL | Max | Min | e = | Loadcell | Emax |
|----------|--------|---------|-------|------------|--------|
| GFK 60M | 60 kg | 0.40 kg | 20 g | HBM PW24C3 | 100 kg |
| GFK 60M | 60 kg | 0.40 kg | 20 g | Zemic L6G | 100 kg |
| GFK 150M | 150 kg | 1 kg | 50 g | HBM PW24C3 | 200 kg |
| GFK 150M | 150 kg | 1 kg | 50 g | Zemic L6G | 200 kg |
| GFK 300M | 300 kg | 2 kg | 100 g | Zemic L6G | 500 kg |

Any compatible load cell may be used providing the following conditions are met:

- There is a respective OIML Certificate of Conformity (R60) issued for the load cell.
- The certificate contains the load cell types and the necessary load cell data required for the manufacturer's declaration of compatibility of modules and any particular installation requirements. A load cell marked NH is allowed only if humidity testing to R76 has been conducted on this load cell.
- The compatibility of the load cells and indicator is established by the manufacturer by means of the compatibility of modules calculation.

Technical data:

| Power supply | 230VAC, 50/60 Hz | |
|--|-------------------------|--|
| | 6V rechargeable battery | |
| Maximum number of scale intervals | 3000 | |
| Load cell excitation voltage | 5 Vdc | |
| Minimum load cell impedance | 87 Ω | |
| Maximum load cell impedance | 1120 Ω | |
| Minimum input voltage per verification scale interval | 1.5 μV | |
| Measuring range minimum voltage | 4.5 mV | |
| Measuring range maximum voltage | 45.0 mV | |
| Fraction of maximum permissible error | $P_{ind} = 0.5$ | |
| Operating temperature range | -10 °C to +40 °C | |
| Load cell connection | 6-wire shielded | |
| Load cell cable max length (junction box to indicator) | 172 m/mm ² | |

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.