



United Kingdom of Great Britain and Northern Ireland

OIML Certificate No R76/1992-GB1-10.08

OIML CERTIFICATE OF CONFORMITY

Issuing authority Name:

Person responsible:

Address:

Applicant Name:

Address:

National Weights and Measures Laboratory (Part of the National Measurement Office) Stanton Avenue Teddington Middlesex TW11 0JZ United Kingdom Paul Dixon - Product Certification Manager Ryco Equipment Inc. 6810 220th Street SW Mountlake Terrace Bellingham WA 98043

Identification of the certified pattern:

#821G, non-automatic weighing instrument

Further characteristics see page 2

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML:	R76
Edition:	1992 (E)
Accuracy class:	III

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

USA

OIML Certificate No R76/1992-GB1-10.08

The conformity was established by tests described in the associated:

TR 584

P00479

NWML Test reports: Pattern evaluation report: having 39 pages having 13 pages

The issuing authority

Mr P R Dixon

Date: 28 September 2010 Ref: TS1201/0005 The CIML member

Mr P Mason

Characteristics: The instrument is a #821G, Class III or IIII, mains or battery-operated, selfindicating, single-interval, non-automatic weighing instrument.

It consists of a #821G indicator connected to a weighing platform.

Main features:

- LCD display
- Operator's keypad (numerical and function keys)
- Stainless steel casing

Devices:

- Initial zero-setting ($\leq 20\%$ Max)
- Semi-automatic zero setting ($\leq 4\%$ Max)
- Zero-tracking ($\leq 4\%$ Max)
- Semi-automatic subtractive tare weighing (T = -20% Max)
- Zero-indicator
- Determination of stability of equilibrium
- Net indicator
- Units change (kg or g)
- Weight bar
- Totalisation

Technical data:

Power supply	100-240 VAC
Maximum number of scale intervals	5,000
Load cell excitation voltage	5 Vdc
Minimum load cell impedance	350 Ω
Maximum load cell impedance	1000 Ω
Minimum input voltage per verification scale interval	1 μV
Measuring range minimum voltage	0 mV
Measuring range maximum voltage	36 mV
Fraction of maximum permissible error	$P_{ind} = 0.5$
Operating temperature range	- 10 °C to + 50 °C
Load cell cable (from indicator to load cell junction box) - Maximum length	2.5 m (4-wire configuration) 278 m/mm ² (6-wire configuration)

Load cell:

Any compatible load cell may be used providing the following conditions are met:

- There is a respective OIML Certificate of Conformity (R60) issued for the load cell
- The certificate contains the load cell types and the necessary load cell data required for the manufacturer's declaration of compatibility of modules and any particular installation requirements. A load cell marked NH is allowed only if humidity testing to R76 has been conducted on this load cell
- The compatibility of the load cells and indicator is established by the manufacturer by means of the compatibility of modules calculation.

Interfaces:

4 or 6-wire load cell connection (1 channel only allowed under this certificate)

Certificate History

ISSUE NO.	DATE	DESCRIPTION
R76/1992-GB1-10.08	28 September 2010	Certificate first issued.
-	-	No revisions have been issued.

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.