

Member State of OIML Japan



OIML Certificate No R76/1992-JP1-06.01

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name:National Institute of Advanced Industrial Science and TechnologyAddress:1-3-1, Kasumigaseki, Chiyoda-ku, TOKYO, 100-8921, JapanPerson responsible:Hiroyuki Yoshikawa

Applicant

Name:TANITA CORPORATIONAddress:14-2, 1chome, Maeno-cho, Itabashi-ku, Tokyo, 174-8630, Japan

Manufacturer of the certified pattern

Name:	TANITA AKITA CORPORATION
Address:	28-1, Shimotamogizoe, Horiminai, Aza, Daisen-shi, Akita-ken,
	014-0113, Japan

Identification of the certified pattern:

Type PW-630MA Further characteristics see page 2

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML):

R76-1, edition 1992, including amendment1 (1994), for accuracy class

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.



Member State of OIML Japan



OIML Certificate No R76/1992-JP1-06.01

The conformity was established by tests described in the associated test report N° 05-03/R76 (56 pages).

The Issuing Authority

湖向砂道満己 産業時編編向研 院稲理事読空 フくばセンター Dr. H. Yoshikawal憲字辞事用

President

Jan 17 2006

The CIML member

Cy. Miki

Dr. Y. Miki

Jan 17 2006

Characteristics:

Туре	PW-630MA
Accuracy class	
Max	300 kg
e ≧	0.1 kg
$d \geq$	0.01 kg
n	3000
Min	2 kg
Tare-balancing	T≤−150 kg
range	
Temperature	0 °C / 35 °C
range	0 0733 0

Important note: Apart from the mention of certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.