


OIML Member State
Japan



OIML Certificate N°
R76/1992-JP1-06.04

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: National Metrology Institute of Japan /National Institute of
Advanced Industrial Science and Technology (NMIJ/AIST)
Address: AIST Tsukuba Central 3-9
Tsukuba Ibaraki 305-8563, Japan
Person responsible: Dr. Hiroyuki Yoshikawa, President of AIST

Applicant

Name: YAMATO SCALE CO., LTD.
Address: 5-22 SAENBA-CHO, AKASHI, HYOGO, 673-8688, JAPAN

Manufacturer of the certified type

Name: YAMATO SCALE CO., LTD.
Address: 5-22 SAENBA-CHO, AKASHI, HYOGO, 673-8688, JAPAN

Identification of the certified type:

Type UDS-1V-WP, UDS-1V II -WP, TCW-WP


Further characteristics see page 2

This certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76-1, Edition 1992, including amendment1 (1994),
for accuracy class: **III**

This certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This certificate does not bestow any form of legal international approval.



OIML Member State
Japan



OIML Certificate N°
R76/1992-JP1-06.04

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° 06-03-A/R76 that includes 28 pages;

N° 06-03-B/R76 that includes 46 pages;

N° 06-03-C/R76 that includes 18 pages.

The Issuing Authority

The OIML Member

NMIJ/AIST



Dr. H. Yoshikawa

Dr. Y. Miki

President of AIST

2006-07-12

2006-07-12

Characteristics:

Type	UDS-1V-WP, UDS-1V II-WP	UDS-1V-WP, UDS-1V II-WP, TCW-WP	
Class	III		
Max (Max ₁ / Max ₂)	1.5 kg / 3 kg	3 kg / 6 kg	7.5 kg / 15 kg
e= (e ₁ / e ₂)	1 g / 2 g	2 g / 5 g	5 g / 10 g
n= (n ₁ / n ₂)	1500 / 1500	1500 / 1200	1500 / 1500
Min	20 g	40 g	100 g
Tare-balancing range	T ≤ - 1.495 kg	T ≤ - 2.99 kg	T ≤ - 7.475 kg
Temperature range	- 10 °C / 40 °C		

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.