



OIML Member State
Japan



OIML Certificate N^o
R76/1992-JP1-08.01

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: National Metrology Institute of Japan /National Institute of Advanced Industrial Science and Technology (NMIJ/AIST)
Address: AIST Tsukuba Central 3-9
Tsukuba Ibaraki 305-8563, Japan
Person responsible: Dr. Hiroyuki Yoshikawa, President of AIST

Applicant

Name: A&D Company, Limited
Address: 3-23-14, Higashi-ikebukuro, Toshima-ku, Tokyo 170-0013,
JAPAN

Manufacturer of the certified type

Name: A&D ELECTRONICS(Shen Zhen)CO.,LTD.
Address: Jiang Bian Cun Di San Gong Ye Qu, Song Gang Zhen, Bao An Qu,
Shen Zhen Shi, Guang Dong Sheng, CHINA

Identification of the certified type:

Type SJ-.../ SJ-12K
Further characteristics see page 2

This certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76-1, Edition 1992, including amendment1 (1994),
for accuracy class: **III**

This certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This certificate does not bestow any form of legal international approval.



OIML Member State
Japan


OIML Certificate N°
R76/1992-JP1-08.01

The conformity was established by the results of tests and examinations provided in the associated Evaluation Report N° R76/1992-JP1-08.01-ER.

The Issuing Authority
NMIJ/AIST

The OIML Member

Dr. H. Yoshikawa
President of AIST
2008-04-07



Dr. Y. Miki
2008-04-07

Characteristics:

Type	SJ-1000	SJ-2000	SJ-5000	SJ-12K
Class	III			
Max (Max ₁ /Max ₂)	1000 g	2000 g	5000 g	10 kg / 12 kg
e (e ₁ /e ₂)	1 g	2 g	5 g	10 g / 20 g
n (n ₁ /n ₂)	1000	1000	1000	1000 / 600
Min	20 g	40 g	100 g	200 g
Maximum Tare-balancing effect	T = - 1000 g	T = - 2000 g	T = - 5000 g	T = - 9.99 kg
Temperature range	- 10 °C / + 40 °C			

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Evaluation Report is not permitted, although either may be reproduced in full.