

OIML Certificate N° R76/1992-NL1-05.01

Project number 410724

Page 1 of 2

The Netherlands

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name:

NMi Certin B.V.

Address:

Hugo de Grootplein 1, Dordrecht

Person responsible:

Ing. C. Oosterman

Applicant

Name:

Snowrex International Co., Ltd.

Address:

2F, N°. 9 Lane 50, Section 3

Nan-Kang Road

Taipei

Taiwan R.O.C.

Manufacturer of the certified type

Name:

Snowrex International Co., Ltd.

Address:

2F, N°. 9 Lane 50, Section 3

Nan-Kang Road

Taipei

Taiwan R.O.C.

Identification of the certified type

Type: NE-Series

 $3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$

e ≥ 1 q

 $n \le 3000$ divisions (per partial weighing range)

maximum of 2 partial weighing ranges

Temperature range 0°C / 40°C

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T6692 and the appertaining documentation folder), with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76

Edition 1992

for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of the instrument covered by the relevant OIML International Recommendation identified above.

Nederlands Meetinstituut Hugo de Grootplein 1 3314 EG Dordrecht Telephone +31 78 6332332

Telefax +31 78 6332309

NMi B.V.

(Chamber of Commerce no.27.228.701)

Subsidiary companies:

NMi Van Swinden Laboratorium B.V. (27228703) NMi Certin B.V. (27.233.418) Verispect B.V. (27.228.700)

This document is issued under the provision that NMi. B.V. nor its subsidiary companies accept any liability.

Reproduction of the complete document is allowed. Parts of the document may only be reproduced after written permission



OIML Certificate N° R76/1992-NL1-05.01

Project number 410724

Page 2 of 2

OIML Member State

The Netherlands

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° R76/1992-NL1-03.19 that includes 54 pages; N° R76/1992-NL1-03.43 that includes 13 pages; N° R76/1992-NL1-05.01 that includes 15 pages.

The Issuing Authority Ing. C. Ogsterman

Manager Product Certification

20 January 2005

The CIML Member Dr. ir. C.J. van Mullem

20 January 2005

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.