



## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: NMI Certin B.V.  
Address: Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
Person responsible: Ing. C. Oosterman

### Applicant

Name: Kingship Weighing Machine Corp.  
Address: 739, Renhua Road  
Dali City, Taichung 412  
Taiwan R.O.C.

### Manufacturer of the certified type

Name: Kingship Weighing Machine Corp.  
Address: 739, Renhua Road  
Dali City, Taichung 412  
Taiwan R.O.C.

### Identification of certified type

Non-automatic weighing instrument  
Type: GEW... / MEP... / MEPL...

$6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$   
 $e \geq 2 \text{ g}$   
 $n \leq 3000$  divisions (per partial weighing range)  
Maximum of two partial weighing ranges  
 $T \leq -50\% \text{ Max}$  or  $T \leq -\text{Max}$ ,  
Temperature range  $-10^\circ\text{C} / +40^\circ\text{C}$



Nederlands Meetinstituut

**OIML Member state**

The Netherlands

**OIML Certificate N° R76/1992-NL1-06.05 Rev. 1**

Project number 504469

Page 2 of 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T6906 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76**

Edition 1992 (E)

for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

N° R76/1992-NL1-06.05A that includes 26 pages;

N° R76/1992-NL1-06.05B1 that includes 39 pages;

N° R76/1992-NL1-06.05B2 that includes 21 pages.

This revision replaces the earlier version.

The Issuing Authority  
Ing. C. Oosterman  
Manager Product Certification

28 February 2006

\*  
\* \*

**Important note:** Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.