



OIML Member state
The Netherlands

OIML Certificate N° R76/1992-NL1-07.01

Project number 610791

Page 1 of 2

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Avery Berkel
Address: Foundry Lane
Smethwick, Warley
West Midlands B66 2LP
United Kingdom

Manufacturer of the certified type

Name: Avery Berkel
Address: Foundry Lane
Smethwick, Warley
West Midlands B66 2LP
United Kingdom

Identification of certified type

Non-automatic weighing instrument
Type : FX, MC, B, G, BK, HL or S series

$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$

$e \geq 1 \text{ g}$

$n \leq 7500$ divisions (single interval instrument)

$n \leq 3000$ division (per partial weighing range)

maximum of two weighing ranges



OIML Member state

The Netherlands

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T5532 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76

Edition 1992 (E)

for accuracy class III

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° R76/1992-NL-99.06, that includes 54 pages;

N°. R76/1992-NL-99.14, that includes 30 pages;

N° R76/1992-NL-01.38, that includes 14 pages;

N° R76/1992-NL1-03.28a that includes 22 pages;

N° R76/1992-NL1-04.18 that includes 11 pages;

N° R76/1992-NL1-07.01a that includes 13 pages;

N° R76/1992-NL1-07.01b that includes 12 pages.

The Issuing Authority

Ing. C. Oosterman

Manager Product Certification

22 January 2007

*

**

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.