



OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Rice Lake Weighing Systems
Address: 230 West Coleman Street
Rice Lake, WI 54868
United States of America

Manufacturer of the certified type

Name: Rice Lake Weighing Systems
Address: 230 West Coleman Street
Rice Lake, WI 54868
United States of America

Identification of certified type

Non-automatic weighing instrument
Type: 420, 420 Plus, 420 HE
 $3 \text{ kg} \leq \text{Max} \leq 100 \text{ t}$
 $1 \text{ g} \leq e \leq 100 \text{ kg}$
 $n \leq 3000$ divisions for class II instruments
 $n \leq 1000$ divisions for class III instruments
Temperature range $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T5692 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76

Edition 1992 (E)

for accuracy class III or III



OIML Member state

The Netherlands

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° 503889A, that includes 42 pages;
N° 503889B, that includes 16 pages;
N° R60/1991-NL-98.02, that includes 38 pages;
N° R60/1991-NL-96.03, that includes 37 pages;
N° R60/1991-NL-96.05, that includes 35 pages;
N° R60/2000-NL1-04.02A, that includes 38 pages;
N° R60/2000-NL1-04.02B, that includes 8 pages;
N° R60/2000-NL1-04.02C, that includes 37 pages;
N° R60/1991-NL-97.03A, that includes 37 pages;
N° R60/1991-NL-97.03B, that includes 35 pages;
N° R60/2000-NL-00.10, that includes 35 pages;
N° R60/2000-NL1-03.11, that includes 37 pages;
N° R60/2000-NL1-01.18, that includes 40 pages;
N° R60/2000-NL1-02.01, that includes 39 pages.
N° R60/1991-GB-94.09 with test report 12-93, that includes 26 pages;
N° R60/1991-GB-95.01 with test report 10-93, that includes 26 pages;
N° R60/1991-GB-95.01 with test report 11-93, that includes 26 pages;
N° R60/1991-GB-95.02 with test report 09-93, that includes 26 pages;
N° R60/1991-GB-95.03 with test report 08-93, that includes 26 pages;
N° R60/1991-GB-95.04 with test report 16-93, that includes 26 pages;
N° R60/1991-GB-95.05 with test report 05-93, that includes 26 pages;
N° R60/1991-GB-95.06 with test report 21-93, that includes 26 pages;
N° R60/1991-GB-95.11 with test report 14-93, that includes 26 pages;
N° R60/1991-GB-95.13 with test report 06-93, that includes 26 pages;
N° R60/1991-GB-95.14 with test report 15-93, that includes 26 pages;
N° R60/1991-GB-00.07 with test report AB02253, that includes 21 pages.

The Issuing Authority

Ing. C. Oesterman

Manager Product Certification

12 February 2007

*
* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.