



OIML Member state
The Netherlands

OIML Certificate N° R76/1992-NL1-08.26

Project number 805386

Page 1 of 2

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Mettler-Toledo Inc.
Address: 1150 Dearborn Drive
Worthington, OH 43085-666712
United States of America

Manufacturer of the certified type

Name: Mettler-Toledo Inc.
Address: 1150 Dearborn Drive
Worthington, OH 43085-666712
United States of America

Identification of certified type

Non-automatic weighing instrument
Type: IND310

$6000 \text{ kg} \leq \text{Max} \leq 60000 \text{ kg}$
 $e \geq 2 \text{ kg}$
 $n \leq 3000 \text{ divisions}$

Temperature range -10 °C / +40 °C



OIML Member state

The Netherlands

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T2206 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76

Edition 1992 (E)

for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° 400460A that includes 45 pages;

N° 400460B that includes 17 pages;

N° 407976 that includes 6 pages;

N° 805386 that includes 12 pages;

N° R60/1991-NL-97.31 that includes 37 pages.

The Issuing Authority

Ing. C. Oosterman

Manager Product Certification

4 July 2008

*

**

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.