



OIML Member state
The Netherlands

OIML Certificate N° R76/1992-NL1-08.28

Project number 805389

Page 1 of 2

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Charder Electronic Co. Ltd.
Address: 103 Kuo Chung Road
Dah Li City, Taichung Hsien 41262
Taiwan R.O.C.

Manufacturer of the certified type

Name: Charder Electronic Co. Ltd.
Address: 103 Kuo Chung Road
Dah Li City, Taichung Hsien 41262
Taiwan R.O.C.

Identification of certified type

Non-automatic weighing instrument
Type: MHS2200/ MHS2400/ MHS2500/ MHS2600

Max \leq 400 kg
e \geq 0.1 kg
n \leq 3000 divisions

T \leq - Max

Temperature range 0°C / +40°C



OIML Member state

The Netherlands

OIML Certificate N° R76/1992-NL1-08.28

Project number 805389

Page 2 of 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T7001 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76

Edition 1992 (E)

for accuracy class III

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° R76/1992-NL1-06.26A that includes 40 pages;

N° R76/1992-NL1-06.26B that includes 13 pages;

N° R76/1992-NL1-08.28 that includes 7 pages.

The Issuing Authority
Ing. C. Oosterman
Manager Product Certification

17 July 2008

*
* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.