

OIML Member state

The Netherlands

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.

Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

Person responsible: Ing. C. Oosterman

Applicant

Name: Taiwan Scale Mfg. Co., Ltd.

Address: 282, sec. 3, Hoping w. Road
Taipei
Taiwan**Manufacturer of the certified type**

Name: Taiwan Scale Mfg. Co., Ltd.

Address: 282, sec. 3, Hoping w. Road
Taipei
Taiwan**Identification of certified type**

Non-automatic weighing instrument

Type: JP / JSP / JPP / JW / JWP

 $3 \text{ kg} \leq \text{Max} \leq 45 \text{ kg}$ $e \geq 1 \text{ g}$ $n \leq 3000$ divisions (per partial weighing range)

maximum of 2 partial weighing ranges

 $T \leq - \text{Max}$ Temperature range $-10^{\circ}\text{C} / +40^{\circ}\text{C}$ **NMI Certin BV**Hugo de Grootplein 1, 3314 EG Dordrecht
PO Box 394, 3300 AJ Dordrecht, NL
T +31 78 6332332
F +31 78 6332309
certin@nmi.nl
www.nmi.nlThis document is issued under the provision
that no liability is accepted and that the
applicant shall indemnify third-party liability.Reproduction of the complete document is
permitted.The notification of NMI Certin as Issuing
Authority can be verified at www.oiml.org.



OIML Certificate N° R76/1992-NL1-09.04 Revision 1

OIML Member state

The Netherlands

Project number 801522

Page 2 of 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T7441 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76

Edition 1992 (E)

for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° R76/1992-NL1-07.35A that includes 53 pages;

N° R76/1992-NL1-07.35B that includes 14 pages

N° R76/1992-NL1-08.03A that includes 14 pages

N° R76/1992-NL1-08.03B that includes 12 pages

N° R76/1992-NL1-08.27 that includes 16 pages

N° R76/1992-NL1-09.04A that includes 24 pages

N° R76/1992-NL1-09.04B that includes 16 pages.

The Issuing Authority NL1
NMI Certin, 26 March 2009


C. Oosterman
Head Certification Board

*
* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.