

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: NMI Certin B.V.

Address: Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands

### Applicant

Name: Precisa Gravimetrics AG

Address: Moosmattstrasse 32  
8953 Dietikon  
Switzerland

### Manufacturer of the certified type

Name: Precisa Gravimetrics AG

Address: Moosmattstrasse 32  
8953 Dietikon  
Switzerland

### Identification of certified type

Non-automatic weighing instrument  
Type: 490 series

Max  $\leq$  60.1 kg  
e  $\geq$  1 g  
d  $\geq$  0.1 g  
n  $\leq$  30100 divisions  
T  $\leq$  -Max

Temperature range +10 °C / +30 °C

**OIML Member state**

The Netherlands

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T7542 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76**

Edition 1992 (E)

for accuracy class **II**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° R76/1992-NL1-09.13 that includes 52 pages.

This revision OIML Certificate of Conformity replaces the earlier version.

The Issuing Authority NL1

NMi Certin, 27 August 2009



C. Oosterman

Head Certification Board

\*

\*\*

**Important note:** Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.