

**OIML Member State**  
The Netherlands

Number R76/1992-NL1-10.16 revision 1  
Project number SO15203176  
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant	A&D Instruments Ltd. 24 Blacklands Way, Abingdon Business Park, Abingdon, Oxfordshire, OX14 1DY United Kingdom
Manufacturer	A&D Instruments Ltd. 24 Blacklands Way, Abingdon Business Park, Abingdon, Oxfordshire, OX14 1DY United Kingdom  A&D SCALES CO., LTD. 191, Inseok-ro, Deoksan-myeon, Jincheon-gun, Chungcheongbuk-do, 365-842, Korea  KENSEI KOGYO Co., Ltd. 4210-15, Takasai, Shimotsuma-shi, Ibaraki, 304-0031, Japan  A&D Company, Limited 3-23-14 Higashi-Ikebukuro, Toshima-ku, Tokyo, 170-0013, Japan
Identification of the certified type	<b>A Non-automatic weighing instrument</b> Type : FZ-i series
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R 76** - Edition 1992 for accuracy class **II**

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**  
29 September 2015



C. Oosterman  
Head Certification Board

NMi Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see [www.nmi.nl](http://www.nmi.nl)).





# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R76/1992-NL1-10.16 revision 1  
Project number SO15203176  
Page 2 of 2

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified.  
This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report:

- Number R76/1992-NL1-10.16 dated 17 May 2010 that includes 49 pages.

## Characteristics of the non-automatic weighing instrument:

Accuracy class	II
Maximum capacity	$122 \text{ g} \leq \text{Max} \leq 3200 \text{ g}$
Verification scale interval	$10 \text{ mg} \leq e < 1 \text{ g}$
Maximum number of scale intervals	$n \leq 32000$ divisions
Temperature range	+10 °C / +30 °C

## Revision History

This revision replaces the previous version.

Revision	Date	Change(s)
Initial	17 May 2010	-
1	29 September 2015	Added multiple manufacturer names.