

**OIML Member State**  
The Netherlands

Number R76/1992-NL1-11.07  
Project number 11200180  
Page 1 of 2

Issuing authority	NMi Certin B.V. Person responsible: C. Oosterman
Applicant	Mettler-Toledo (Changzhou) Measurement Technology Ltd. 111, West Taihu Road, XinBei District, Changzhou, Jiangsu, 213125, Peoples Republic of China
Manufacturer	Mettler-Toledo (Changzhou) Measurement Technology Ltd. 111, West Taihu Road, XinBei District, Changzhou, Jiangsu, 213125, Peoples Republic of China
Identification of the certified type	<b>A Non-automatic weighing instrument</b> Type : PCA320, PCA220 series
Characteristics	See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML R76** - Edition 1992 for accuracy class **(III)**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

*Important note:* Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**  
2 May 2011

  
C. Oosterman  
Head Certification Board

NMi Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
the Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org)

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see [www.nmi.nl](http://www.nmi.nl)).





# OIML Certificate of Conformity

**OIML Member State**  
The Netherlands

Number R76/1992-NL1-11.07  
Project number 11200180  
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- Number R76/1992-NL1-10.03, dated 23 November 2010 that includes 50 pages.
- Number NMI-11200180-01, dated 28 April 2011 that includes 15 pages;
- Number NMI-11200180-02, dated 24 April 2011 that includes 28 pages;

## Characteristics of the non-automatic weighing instrument:

Maximum capacity	:	$100 \text{ kg} \leq \text{Max} \leq 100000 \text{ kg}$
Verification scale interval	:	$e \geq 50 \text{ g}$
Maximum number of scale intervals	:	$n \leq 3000 \text{ divisions}$
Temperature range	:	$-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$
Tare	:	$T \leq -\text{Max}$