

OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R76/1992-NL1-12.15 Project number 12200035 Page 1 of 2

NMi Certin B.V. Issuing authority

Person responsible: C. Oosterm

Applicant A&D Company, Limited

3-23-14 Higashi Ikebukuro

Toshima Ku Tokyo 170-0013

Japan

Manufacturer A&D SCALES CO., LTD.

> 162-4, Insan-ni, Deogsan-myeon, Jincheon-gun, Chungcheongbug-do

356-842 KOREA

Identification of the

A Non-automatic weighing instrument

certified type

Type

Characteristics See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation International Organization of Legal Metrology (OIML):

OIML R76 - Edition 1992 for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

10 April 2012

C. Oosterman Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht the Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).







OIML Certificate of Conformity

OIML Member State The Netherlands

Number R76/1992-NL1-12.15 Project number 12200035 Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

Number NMi-12200035-01 dated 10 April 2012 that includes 54 pages; Number NMi-12200035-02 dated 10 April 2012 that includes 21 pages.

Characteristics of the non-automatic weighing instrument:

Maximum capacity $1 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$

Verification scale interval e ≥ 1 g

Maximum number of scale intervals $n \le 1000$ divisions

(per partial weighing range)

Maximum partial weighing ranges

-10 °C / +40 °C Temperature range

Tare T ≤ -100% for single interval instruments