

OIML Certificate of Conformity

OIML Member State

The Netherlands

Number R76/1992-NL1-13.16 Project number 13200262 Page 1 of 2

Issuing authority NMi Certin B.V.

Person responsible: C. Oosterma

Shimadzu Corporation **Applicant**

1, Nishinokyo-Kuwabaracho

Nakagyo-ku, Kyoto 604

Japan

Manufacturer Shimadzu Corporation

> 1, Nishinokyo-Kuwabaracho Nakagyo-ku, Kyoto 604

Identification of the

certified type

A Non-automatic weighing instrument

Type

Characteristics See next page

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R76 - Edition 1992 for accuracy class (1) and/or (11)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

NMi Certin B.V.,

6 June 2013

C. Oosterman Head Certification Board

NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht the Netherlands T+31 78 6332332 certin@nmi.nl www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).







OIML Certificate of Conformity

OIML Member StateThe Netherlands

Number R76/1992-NL1-13.16 Project number 13200262 Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

Number NMi-13200262-01 dated 23 May 2013 that includes 16 pages; Number R76/1992-NL1-10.23a dated 19 July 2010 that includes 38 pages; Number R76/1992-NL1-10.23b dated 19 July 2010 that includes 16 pages.

Characteristics of the non-automatic weighing instrument:

Maximum capacity : $64 \text{ g} \le \text{Max} \le 4200 \text{ g} \text{ or } 320 \text{ ct} \le \text{Max} \le 21000 \text{ ct}$

Verification scale interval : $0.01 \text{ g} \le e < 1 \text{ g}$ or $0.01 \text{ ct} \le e < 5 \text{ ct}$

e = 10 d

Maximum number of scale intervals : $n \le 62000$ divisions

Temperature range : $+10 \,^{\circ}\text{C} / +30 \,^{\circ}\text{C}$