



**FEDERAL AGENCY FOR TECHNICAL REGULATION  
AND METROLOGY OF RUSSIA**

OIML Member State  
The Russian Federation

OIML Certificate №  
R 76/1992-RU1-06.01

**OIML CERTIFICATE OF CONFORMITY**

**Issuing authority**

Name: VNIIMS  
Address: 46, Ozernaya str, G-361, Moscow, 119361,  
the Russian Federation  
Person responsible: Dr. S.A. Kononogov

**Applicant**

Name: AVITEC-PLUS Ltd  
Address: 122, Malisheva str. Yekaterinburg,  
Sverdlovsk region, 620078  
The Russian Federation

**Manufacturer of the certified pattern is the Applicant**

Identification of the certified pattern: ABP-B-CD

This certificate attests the conformity of the above-mentioned pattern (represented by the sample identified in the associated test reports) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R 76  
for accuracy class III

edition 1992 (E)  
including Amendment 1 (1994)

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in associated test report number 401-06/04 which includes 53 pages.

**The issuing authority**

**VNIIMS**

**Dr. S.A. Kononogov**  
**Director**  
**Date: 23.04.2006**



**The OIML member**

**Dr. L.K. Issaev**  
**Professor**  
**Date: 23.04.2006**

Characteristics:

Type	ABP-B-CD			
Model	ABP-B-CD-60	ABP-B-CD-100	ABP-B-CD-150	ABP-B-CD-200
Accuracy class	III			
Max	60	100	150	200
Min	6	8	8	16
e = d	20	50	50	100
n ≤	3000	2000	3000	2000
Temperature range °C	- 30...+ 50			

Important note: Apart from the mention of the certificates reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.