



Member State
Switzerland

OIML Certificate
No R76_2006-CH1-09.01

OIML CERTIFICATE OF CONFORMITY

Issuing authority

| | |
|---------------------------|---|
| <i>Name</i> | Federal Office of Metrology METAS Certification Body METAS-Cert |
| <i>Address</i> | Lindenweg 50 CH-3003 Bern-Wabern |
| <i>Person responsible</i> | Jürg Ramseyer, Head of METAS-Cert |

Applicant

| | |
|----------------|--|
| <i>Name</i> | METTLER-TOLEDO AG LabTec |
| <i>Address</i> | Im Langacher 8606 Greifensee Switzerland |

Manufacturer The manufacturer of the certified pattern is the Applicant

Identification of the certified pattern

Non-automatic weighing instrument

Type **NewClassic MF**

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):



R 76-1, edition 2006

for accuracy class(es) (I), (II)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.



**OIML Certificate
No R76_2006-CH1-09.01**

Metrological characteristics

| Accuracy class | I | II | |
|----------------------|--------------------|-------------------|----------------------|
| | | Small housing | Large housing |
| Max | ≤ 10 200 g | 120 g to 10 200 g | 10 200 g to 64 200 g |
| e | ≥ 1 mg | ≥ 10 mg | ≥ 1 g |
| d | ≥ 0.1 mg | ≥ 1 mg | ≥ 0.1 g |
| n | ≤ 420 000 | ≤ 82 000 | ≤ 64 200 |
| Tare-balancing range | Until 100 % of Max | | |
| Temperature range | 10 °C / 30 °C | | 5 °C / 40 °C |

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

| No | Including pages | No | Including pages |
|-----------|-----------------|-----------|-----------------|
| 511-00144 | 28 | 511-00348 | 22 |
| 511-00152 | 32 | 511-00351 | 17 |
| 511-00306 | 32 | 511-00354 | 28 |
| 511-00308 | 18 | 511-00357 | 14 |
| 511-00339 | 18 | 511-00361 | 28 |
| 511-00342 | 28 | 511-00364 | 28 |
| 511-00345 | 28 | 511-00441 | 19 |
| 511-00346 | 22 | | |

The Issuing Authority

The OIML Member

Federal Office of Metrology METAS

Jürg Ramseyer, Head of METAS-Cert

Dr. Philippe Richard, Vice Director

CH-3003 Bern-Wabern, 7 October 2009

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and the associated Test Report is not permitted, although either may be reproduced in full.