

Member State Switzerland

OIML Certificate No R76/2006-CH1-13.01 Rev. 1

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name

Federal Institute of Metrology METAS

Address

Lindenweg 50, 3003 Bern-Wabern, Switzerland

Person responsible

Gulian Couvreur, Head of METAS-Cert

Applicant

Name

seca gmbh & co. kg

Address

Hammer Steindamm 9-25, 22089 Hamburg, Deutschland

Manufacturer

The manufacturer of the certified pattern is the Applicant

Identification of the certified pattern

Non-automatic electro-mechanic infantile weighing

instrument

Type

BIS04...

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):



R 76-1, edition 2006

for accuracy class(es)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

OIML Certificate No R76/2006-CH1-13.01 Rev. 1

Metrological characteristics

Max	5 000 g	15 000 g
е	2 g	5 g
n	2500	3000
Temperature range	10 °C / 40 °C	

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

No / Including pages

218-01191 / 16 511-01404 / 55

CH-3003 Bern-Wabern, 9 September 2014

The Issuing Authority

Federal Institute of Metrology METAS

Gulian Couvreur, Head of METAS-Cert

The CIML Member

Dr. Philippe Richard, Deputy Director

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and the associated Test Report is not permitted, although either may be reproduced in full.