

Schweizerische Eidgenossenschaft Confédération suisse Confederazione Svizzera Confederaziun svizra

Member State Switzerland

OIML Certificate No R76/2006-CH1-15.01

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name	Federal Institute of Metrology METAS		
Address	Lindenweg 50, 3003 Bern-Wabern, Switzerland		
Person responsible	Gulian Couvreur, Head of METAS-Cert		
Applicant			

Name	IRD Inc. International Road Dynamics
Address	702-43rd Street East, S7K 3T9 Saskatoon, Saskatchewan, Canada
Manufacturer	
Name:	DISA Elektro AG

Kägiswilerstrasse 33, 6060 Sarnen, Switzerland

Identification of the certified pattern

Address:

Non-automatic wheel and axle weighing instrument

Type

SAW...III

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):



R 76-1, edition 2006

for accuracy class(es)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

This document may not be published or forwarded other than in full.

OIML Certificate No R76/2006-CH1-15.01

Metrological characteristics

			For Accuracy Class
Max	10 000 kg to	0 15 000 kg	
е	20 kg to	50 kg	
n	≤ 500		
Temperature range	-20 °C / 40 °C		

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

No / Including pages

133-07081 / 51 133-07082 / 11 218-01230 / 11

CH-3003 Bern-Wabern, 30 January 2015

The Issuing Authority Federal Institute of Metrology METAS Couvreur, Head of METAS-Cert

The CIML Member

Dr. Philippe Richard, Deputy Director

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and the associated Test Report is not permitted, although either may be reproduced in full.