

Member State
P.R.China



OIML Certificate No.
R76/2006-CNI-13.01

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine
Address: NO.9 Madiandonglu, Haidian District, Beijing, 100088 P.R.China

Applicant

Name: Zhongshan Camry Electronic Co., LTD
Address: Baishawan Industrial Park, Qiwan Road East, East District,
Zhongshan, Guangdong, (528403) P.R.China

Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:
Address:

Identification of the certified type

Non-automatic weighing instrument

Type	ACS-15	ACS-30
Accuracy class	III	
Max	15 kg	30 kg
Min	0.1 kg	0.2 kg
e	5 g	10 g
n	3600	
Temperature range(°C)	-10 °C ~ +10 °C	
Humidity range	≤ 85% RH	
Power supply voltage	AC 110 V ~ 240 V, 50-60 Hz; Rechargeable Battery 4V	

The numbers of the associated test reports and the included pages:

NIM-LSme-1202 57 pages

This certificate attests the conformity of the above-mentioned type(represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: 76

Edition: 2006

For accuracy class (if applicable): Ⅲ

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval. The conformity was established by tests described in the associated test reports.

The issuing authority
Office of OIML Affairs

中国秘书处
January 16, 2013

The CIML member

蒲长城

January 16, 2013

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.