Member State P.R.China



OIML Certificate No. R76/2006-CN1-13.04

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine Address: N0.9 Madiandonglu, Haidian District, Beijing, 100088P.R.China

Applicant

Name:JIANGSU HONSTA ELECTRIC MANUFACTURING CO., LTDAddress:Xihu Road 118, Wujin Hi-tech Industry Zone, Changzhou, Jiangsu, (213161) P.R.China

Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:

Address:

Identification of the certified type

Price Computing Scale

Туре	ACS-6-JJ(F902)	ACS-15-JJ(F902)	ACS-30-JJ(F902)
Accuracy class			
Max	6 kg	15 kg	30 kg
Min	0.04 kg	0.1 kg	0.2 kg
e	2 g	5 g	10 g
n	3000	3000	3000
Temperature range(℃)	0 °C ~40 °C		
Humidity range	≤ 85 %RH		
Power supply voltage	AC 210 V ~ 230 V, 50 Hz, Rechargeable battery: 6 V		

The numbers of the associated test reports and the included pages:

NIM-LSmc-1302 57 pages

OIML Certificate No. R76/2006-CN1-13.04

This certificate attests the conformity of the above-mentioned type(represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

 R:
 76

 Edition:
 2006

For accuracy class (if applicable):

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned ,as covered by the relevant OIML International Recommendation(s). This certificate does not bestow any form of legal international approval. The conformity was established by tests described in the associated test reports.



The CIML member

Мау10, 2013

Important note:Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued,partial quotation of the certificate or of the associated test report is not permitted,though they may be reproduced in full.